

Certificate no: Oxygen Optode 4330F_805_157036
FoillID: 2808F

Product: **Oxygen Optode 4330F**
Serial no: 805
Calibration date: 23.04.2019

2-POINT RECALIBRATION, OXYGEN OPTODE

This is to certify that this product has been calibrated and verified using the following reference equipment:

Reference Equipment	Serial
Calibration Bath Model Toho TM-005-P-A	12-K
Fluke 5615 PRT	Serial No. 802054 and Serial No. 849155
Fluke CHUB E-4	Serial No. A7C677
Reference sensor 4330	338

Specification

	O2 Concentration	Air Saturation	Temperature
Range	0-500 μ M	0 - 120%	-5 to +40°C
Accuracy	< $\pm 8\mu$ M or $\pm 5\%$ ¹⁾	$\pm 5\%$ ²⁾	$\pm 0.03^\circ$ C ³⁾
Resolution	< 1 μ M	< 0.4%	0.01°C
Settling Time (63%)	< 8 seconds	< 8 seconds	< 2 seconds

1) Requires salinity compensation for salinity variations > 1mS/cm, and pressure compensation for pressure > 100meter

2) Within calibrated range 0 - 120% / 0 - 30°C

3) Within calibrated range 0 - 36°C for standard version and 0 - 30°C for multipoint version

Calibration points and readings:

	Air Saturated Water	Zero Solution (Na2SO3)
Phase reading (°)	34.93	60.87
Temperature reading (°C)	9.96	21.95
Air Pressure (hPa)	956.00	

Giving these coefficients

Index	0 (Offset)	1 (Slope)
ConcCoef	-7.716535E-01	1.084765E+00

With following settings

Index	0	1	2	3	4	5	6
SVUFoilCoef	0.00000E+00	0.00000E+00	0.00000E+00	0.00000E+00	0.00000E+00	0.00000E+00	0.00000E+00
PhaseCoef	-9.553927E+00	1.164837E+00	0.000000E+00	0.000000E+00			
TempCoef	2.306829E+01	-3.081902E-02	2.868253E-06	-4.139947E-09	0.000000E+00	0.000000E+00	
Salinity (mS/cm)	0.00						

Date:23.04.2019



Sr. Repair Technician