



Scattering Meter Calibration Sheet

10/28/2011

Wavelength: 700

S/N BB2FLVMT-871

Use the following equation to obtain "scaled" output values:

$$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

- **Scale Factor for 700 nm** = 3.252E-06 ($\text{m}^{-1}\text{sr}^{-1}$)/counts
- **Output** = meter reading counts
- **Dark Counts** = 44 counts

Instrument Resolution = 1.0 counts 3.36E-06 ($\text{m}^{-1}\text{sr}^{-1}$)

Definitions:

- **Scale Factor:** Calibration scale factor, $\beta(\theta_c)/\text{counts}$. Refer to User's Guide for derivation.
- **Output:** Measured signal output of the scattering meter.
- **Dark Counts:** Signal obtained by covering detector with black tape and submersing sensor in water.

Instrument Resolution: Standard deviation of 1 minute of collected data.